

COD 11527 4 x 50 mL	COD 11507 2 x 250 mL
STORE AT 15-30°C	
Reagents for measurement of calcium concentration Only for <i>in vitro</i> use in the clinical laboratory	

CALCIUM - MTB



CALCIUM - MTB METHYLTHYMOL BLUE

PRINCIPLE OF THE METHOD

Calcium in the sample reacts with methylthymol blue in alkaline medium forming a coloured complex that can be measured by spectrophotometry. Hydroxyquinoline is included in the reagent to avoid magnesium interference^{1,2}.

CONTENTS

	COD 11527	COD 11507
A. Reagent	2 x 50 mL	1 x 250 mL
B. Reagent	2 x 50 mL	1 x 250 mL
S. Standard	1 x 5 mL	1 x 5 mL

COMPOSITION

- A. Reagent. Potassium cyanide 7.7 mmol/L, ethanalamine 1.5 mol/L.
 B. Reagent. Methylthymol blue 0.1 mmol/L, hydrochloric acid 10 mmol/L, hydroxyquinoline 17 mmol/L.
 S. Calcium/Magnesium Standard. Calcium 10 mg/dL (2.5 mmol/L), magnesium 2 mg/dL. Aqueous primary standard.

STORAGE

Store at 15-30°C.

Reagents and Standard are stable until the expiry date shown on the label when stored tightly closed and if contaminations are prevented during their use.

Reagent B may present aggregates which do not affect its functionality. Mix before use.

Indications of deterioration:

- Reagents: Absorbance of the blank over 0.800 at 610 nm.
- Standard: Presence of particulate material, turbidity.

REAGENT PREPARATION

Standard (S) is provided ready to use.

Working Reagent: Mix equal volumes of Reagent A and Reagent B (Note 1). Mix gently. Stable for 2 days at 2-8°C.

ADDITIONAL EQUIPMENT

- Analyzer, spectrophotometer or photometer able to read at 610 ± 20 nm.

SAMPLES

Serum, heparinized plasma or urine collected by standard procedures.

Calcium in serum or plasma is stable for 10 days at 2-8°C. Anticoagulants other than heparin should not be used.

Collect a 24-hour urine specimen in a bottle containing 10 mL of 50 % (v/v) nitric acid. Stable for 10 days at 2-8°C. Centrifuge or filter and dilute ½ with distilled water before testing.

PROCEDURE

- Pipette into labelled test tubes: (Notes 1, 2)

	Blank	Standard	Sample
Calcium Standard (S)	—	10 µL	—
Sample	—	—	10 µL
Working Reagent	1.0 mL	1.0 mL	1.0 mL

- Mix thoroughly and let stand the tubes for 2 minutes at room temperature.
- Read the absorbance (A) of the Standard and the Sample at 610 nm against the Blank. The colour is stable for at least 1 hour.

CALCULATIONS

The calcium concentration in the sample is calculated using the following general formula:

$$\frac{A_{\text{Sample}}}{A_{\text{Standard}}} \times C_{\text{Standard}} \times \text{Sample dilution factor} = C_{\text{Sample}}$$

If the Calcium Standard provided has been used to calibrate (Note 3):

	Serum and plasma	Urine
$\frac{A_{\text{Sample}}}{A_{\text{Standard}}}$	x 10 = mg/dL calcium	x 20 = mg/dL calcium
	x 2.5 = mmol/L calcium	x 5 = mmol/L calcium

REFERENCE VALUES

Serum and plasma³: 8.6-10.3 mg/dL = 2.15-2.58 mmol/L

Urine³: 100-300 mg/24-h = 2.5-7.5 mmol/24-h

These ranges are given for orientation only; each laboratory should establish its own reference ranges.

QUALITY CONTROL

It is recommended to use the Biochemistry Control Serum level I (cod. 18005, 18009 and 18042) and II (cod. 18007, 18010 and 18043) to verify the performance of the measurement procedure.

Each laboratory should establish its own internal Quality Control scheme and procedures for corrective action if controls do not recover within the acceptable tolerances.

METROLOGICAL CHARACTERISTICS

- Detection limit: 0.6 mg/dL calcium = 0.15 mmol/L calcium.
- Linearity limit: 15 mg/dL calcium = 3.75 mmol/L calcium. For higher values dilute sample 1/2 with distilled water and repeat measurement.
- Repeatability (within run):

Mean calcium concentration	CV	n
11 mg/dL = 2.75 mmol/L	0.7 %	20
13 mg/dL = 3.25 mmol/L	2.6 %	20

- Reproducibility (run to run):

Mean calcium concentration	CV	n
11 mg/dL = 2.75 mmol/L	3.9 %	25
13 mg/dL = 3.25 mmol/L	4.7 %	25

- Sensitivity: 30 mA·dL/mg = 120 mA·L/mmol
- Trueness: Results obtained with this reagent did not show systematic differences when compared with reference reagents (Note 3). Details of the comparison experiments are available on request.
- Interferences: Hemolysis (hemoglobin < 10 g/L) and bilirubin (< 20 mg/dL) do not interfere. Lipemia (triglycerides > 1.25 g/L) interfere. Other drugs and substances may interfere⁴.

These metrological characteristics have been obtained using an analyzer. Results may vary if a different instrument or a manual procedure are used.

DIAGNOSTIC CHARACTERISTICS

Calcium is the most prevalent cation found in the body, distributed in bone (99%), soft tissues and extracellular fluid. Its concentration in plasma is regulated by parathyroid hormone, vitamin D and calcitonin.

Calcium ion is important in the transmission of nerve impulses, in the maintenance of normal muscle contractility, as a cofactor in certain enzyme reactions, and in the coagulation of the blood.

Hypercalcemia can be due to vitamin D intoxication, enhanced renal retention, osteoporosis, sarcoidosis, thyrotoxicosis, hyperparathyroidism, multiple myeloma, idiopathic hypercalcemia of infancy, and carcinoma metastatic to bone^{3,5}.

Elevated calcium concentration in urine is found in nephrolithiasis and metabolic acidosis^{3,5}.

Hypocalcemia may be caused by primary and secondary hypoparathyroidism, pseudohypoparathyroidism, vitamin D deficiency, malnutrition and intestinal malabsorption^{3,5}.

Clinical diagnosis should not be made on the findings of a single test result, but should integrate both clinical and laboratory data.

NOTES

- Contamination of glassware with calcium will affect the test. Use acid-washed glassware or plastic tubes.
- These reagents may be used in several automatic analysers. Instructions for many of them are available on request.
- Calibration with the provided aqueous standard may cause a matrix related bias, specially in some analyzers. In these cases, it is recommended to calibrate using a serum based standard (Biochemistry Calibrator, cod. 18011 and 18044).

BIBLIOGRAPHY

- Gindler M and King JD. Rapid colorimetric determination of calcium in biologic fluids with methylthymol blue. *Am J Clin Path* 1972; 58: 376-382.
- Barnett RN, Skodon SB and Goldberg MH. Performance of kits used for clinical chemical analysis of calcium in serum. *Am J Clin Path* 1973; 59: 836-843.
- Tietz Textbook of Clinical Chemistry and Molecular Diagnostics, 4th ed. Burtis CA, Ashwood ER, Bruns DE. WB Saunders Co, 2005.
- Young DS. Effects of drugs on clinical laboratory tests, 5th ed. AACC Press, 2000.
- Friedman and Young. Effects of disease on clinical laboratory tests, 4th ed. AACC Press, 2001.