

COD 11832 1 x 50 mL	COD 11533 1 x 200 mL	COD 11568 1 x 500 mL	COD 11562 1 x 1 L
STORE AT 2-8°C			
Reagents for measurement of ALT/GPT concentration Only for <i>in vitro</i> use in the clinical laboratory			


ALANINE AMINOTRANSFERASE (ALT/GPT)



ALANINE AMINOTRANSFERASE (ALT/GPT) IFCC

PRINCIPLE OF THE METHOD

Alanine aminotransferase (ALT or GPT) catalyzes the transfer of the amino group from alanine to 2-oxoglutarate, forming pyruvate and glutamate. The catalytic concentration is determined from the rate of decrease of NADH, measured at 340 nm, by means of the lactate dehydrogenase (LDH) coupled reaction^{1,2,3}.



CONTENTS

	COD 11832	COD 11533	COD 11568	COD 11562
A. Reagent	1 x 40 mL	1 x 160 mL	1 x 400 mL	1 x 800 mL
B. Reagent	1 x 10 mL	1 x 40 mL	1 x 100 mL	1 x 200 mL

COMPOSITION

A. Reagent: Tris 150 mmol/L, L-alanine 750 mmol/L, lactate dehydrogenase > 1350 U/L, pH 7.3.

B. Reagent: NADH 1.9 mmol/L, 2-oxoglutarate 75 mmol/L, Sodium hydroxide 148 mmol/L, sodium azide 9.5 g/L.

WARNING: H302: Harmful if swallowed. EUH031: Contact with acids liberates toxic gas. P301+P312: IF SWALLOWED: Call a POISON CENTER or doctor/physician if you feel unwell. P330: Rinse mouth.

For further warnings and precautions, see the product safety data sheet (SDS).

STORAGE

Store at 2-8°C.

Reagents are stable until the expiry date shown on the label when stored tightly closed and if contaminations are prevented during their use.

Indications of deterioration:

- Reagents: Presence of particulate material, turbidity, absorbance of the blank lower than 1.400 at 340 nm (1 cm cuvette).

AUXILIARY REAGENTS

C. Reagent (cod 11666): Pyridoxal phosphate 10 mmol/L. 5 mL.

REAGENT PREPARATION

Working Reagent: Pour the contents of the Reagent B into the Reagent A bottle. Mix gently. Other volumes can be prepared in the proportion: 4 mL Reagent A + 1 mL Reagent B (Note 1). Stable for 1 month at 2-8°C.

Working Reagent with Pyridoxal Phosphate (Note 2): Mix as follows: 10 mL of Working Reagent + 0.1 mL of Reagent C (cod 11666). Stable for 6 days at 2-8°C.

ADDITIONAL EQUIPMENT

- Analyzer, spectrophotometer or photometer with cell holder thermostatable at 30 or 37°C and able to read at 340 nm.
- Cuvettes with 1 cm light path.

SAMPLES

Serum and plasma collected by standard procedures.

Alanine aminotransferase in serum and plasma is stable for 7 days at 2-8°C. Use heparin or EDTA as anticoagulant⁷.

PROCEDURE

- Bring the Working Reagent and the instrument to reaction temperature.
- Pipette into a cuvette: (Note 3)

Reaction temperature	37°C	30°C
Working Reagent	1.0 mL	1.0 mL
Sample	50 µL	100 µL

- Mix and insert the cuvette into the photometer. Start the stopwatch.
- After 1 minute (Note 1), record initial absorbance and at 1 minute intervals thereafter for 3 minutes.
- Calculate the difference between consecutive absorbances, and the average absorbance difference per minute ($\Delta A/\text{min}$).

CALCULATIONS

The ALT/GPT concentration in the sample is calculated using the following general formula:

$$\Delta A/\text{min} \times \frac{V_t \times 10^6}{\epsilon \times l \times V_s} = \text{U/L}$$

The molar absorbance (ϵ) of NADH at 340 nm is 6300, the lightpath (l) is 1 cm, the total reaction volume (V_t) is 1.05 at 37°C and 1.1 at 30°C, the sample volume (V_s) is 0.05 at 37°C and 0.1 at 30°C, and 1 U/L are 0.0166 $\mu\text{kat/L}$. The following formulas are deduced for the calculation of the catalytic concentration:

	37°C	30°C
$\Delta A/\text{min}$	$\times 3333 = \text{U/L}$ $\times 55.55 = \mu\text{kat/L}$	$\times 1746 = \text{U/L}$ $\times 29.1 = \mu\text{kat/L}$

REFERENCE VALUES

Reaction temperature	37°C	30°C
Without pyr-P, up to ³	41 U/L = 0.68 $\mu\text{kat/L}$	29 U/L = 0.48 $\mu\text{kat/L}$
With pyr-P, up to ¹	65 U/L = 1.08 $\mu\text{kat/L}$	35 U/L = 0.58 $\mu\text{kat/L}$

These ranges are given for orientation only; each laboratory should establish its own reference ranges.

QUALITY CONTROL

It is recommended to use the Biochemistry Control Serum level I (cod. 18005, 18009 and 18042) and II (cod. 18007, 18010 and 18043) to verify the performance of the measurement procedure.

Each laboratory should establish its own internal Quality Control scheme and procedures for corrective action if controls do not recover within the acceptable tolerances.

METROLOGICAL CHARACTERISTICS

- Detection limit: 1.6 U/L = 0.027 $\mu\text{kat/L}$
- Linearity limit: 800 U/L = 13.3 $\mu\text{kat/L}$. For higher values dilute sample 1/10 with distilled water and repeat measurement.
- Repeatability (within run):

Mean Concentration	CV	n
43 U/L = 0.72 $\mu\text{kat/L}$	1.8 %	20
192 U/L = 3.2 $\mu\text{kat/L}$	2.8 %	20

- Reproducibility (run to run):

Mean Concentration	CV	n
43 U/L = 0.72 $\mu\text{kat/L}$	5.3 %	25
192 U/L = 3.2 $\mu\text{kat/L}$	2.7 %	25

- Sensitivity: 0.3 $\Delta\text{mA}\cdot\text{L}/\text{U}\cdot\text{min} = 0.00502 \Delta\text{mA}\cdot\text{L}/\mu\text{kat}\cdot\text{min}$
- Trueness: Results obtained with this reagent did not show systematic differences when compared with reference reagents. Details of the comparison experiments are available on request.
- Interferences: Hemolysis (hemoglobin 10 g/L) and bilirubin (20 mg/dL) do not interfere. Lipemia (triglycerides 2 g/L) may affect the results. Other drugs and substances may interfere⁴.

These metrological characteristics have been obtained using an analyzer. Results may vary if a different instrument or a manual procedure are used.

DIAGNOSTIC CHARACTERISTICS

The aminoatransferases catalyze the formation of glutamic acid from 2-oxoglutarate by transfer of amino groups. ALT is normally present in various tissues but its higher concentrations are found in liver and kidney.

The serum concentration of ALT is elevated in hepatitis and other forms of hepatic disease associated with necrosis: infectious mononucleosis, cholestasis, cirrhosis, metastatic carcinoma of the liver, delirium tremens, and after administration of various drugs, such as opiates, salicylates or ampicillin^{5,6}.

Serum ALT concentration can also be elevated in skeletal or cardiac muscle disease^{5,6}.

Clinical diagnosis should not be made on the findings of a single test result, but should integrate both clinical and laboratory data.

NOTES

- The initial absorbance of the reaction mixture may be out of range in some photometers with a low maximum absorbance reading. For these photometers it is recommended to prepare the Working Reagent by mixing in the proportion: 5 mL Reagent A + 1 mL Reagent B.
- The IFCC recommended method specifies the addition of pyridoxal phosphate. The delay time before measurements should then be increased to 2 minutes.
- These reagents may be used in several automatic analysers. Instructions for many of them are available on request.

BIBLIOGRAPHY

- IFCC primary reference procedures for the measurement of catalytic activity concentrations of enzymes at 37°C. Part 4. Reference procedure for the measurement of catalytic concentration of alanine aminotransferase. *Clin Chem Lab Med* 2002; 40: 718-724.
- IFCC reference procedures for measurement of catalytic concentrations of enzymes: corrigendum, notes and useful advice. *Clin Chem Lab Med* 2010; 48: 615-621.
- Gella FJ, Olivella T, Cruz Pastor M, Arenas J, Moreno R, Durban R and Gómez JA. A simple procedure for routine determination of aspartate aminotransferase and alanine aminotransferase with pyridoxal phosphate. *Clin Chim Acta* 1985; 153: 241-247.
- Young DS. Effects of drugs on clinical laboratory tests, 5th ed. AACCPress, 2000.
- Friedman and Young. Effects of disease on clinical laboratory tests, 4th ed. AACCPress, 2001.
- Tietz Textbook of Clinical Chemistry and Molecular Diagnostics, 4th ed. Burtis CA, Ashwood ER, Bruns DE. WB Saunders Co, 2005.
- World Health Organization (WHO). Use of anticoagulants in diagnostic laboratory investigations. Document WHO/DIL/LAB/99.1, Rev.2; 2002.